



# **PHILIPPINE BIDDING DOCUMENTS**

(As Harmonized with Development Partners)

# **Procurement of MEDICAL EQUIPMENT**

**Project Reference No. TF DUE TO NGAS - DOH  
21-011**

Government of the Republic of the Philippines

**“Kaagapay ng Pangulo ang Palaweño  
sa Patuloy na Pagbabago”**

# Table of Contents

<b>Glossary of Acronyms, Terms, and Abbreviations .....</b>	<b>3</b>
<b>Section I. Invitation to Bid.....</b>	<b>6</b>
<b>Section II. Instructions to Bidders.....</b>	<b>10</b>
1. Scope of Bid .....	11
2. Funding Information.....	11
3. Bidding Requirements .....	11
4. Corrupt, Fraudulent, Collusive, and Coercive Practices .....	11
5. Eligible Bidders.....	11
6. Origin of Goods .....	12
7. Subcontracts .....	12
8. Pre-Bid Conference .....	13
9. Clarification and Amendment of Bidding Documents .....	13
10. Documents comprising the Bid: Eligibility and Technical Components .....	13
11. Documents comprising the Bid: Financial Component .....	13
12. Bid Prices .....	14
13. Bid and Payment Currencies .....	15
14. Bid Security .....	15
15. Sealing and Marking of Bids .....	15
16. Deadline for Submission of Bids .....	16
17. Opening and Preliminary Examination of Bids .....	16
18. Domestic Preference .....	16
19. Detailed Evaluation and Comparison of Bids .....	16
20. Post-Qualification .....	17
21. Signing of the Contract .....	17
<b>Section III. Bid Data Sheet .....</b>	<b>18</b>
<b>Section IV. General Conditions of Contract .....</b>	<b>21</b>
1. Scope of Contract .....	22
2. Advance Payment and Terms of Payment .....	22
3. Performance Security .....	22
4. Inspection and Tests .....	22
5. Warranty .....	22
6. Liability of the Supplier .....	23
<b>Section V. Special Conditions of Contract .....</b>	<b>24</b>
<b>Section VI. Schedule of Requirements .....</b>	<b>29</b>
<b>Section VII. Technical Specifications .....</b>	<b>30-39</b>
<b>Section VIII. Checklist of Technical and Financial Documents/Bidding Forms .....</b>	<b>40-59</b>

# *Glossary of Acronyms, Terms, and Abbreviations*

**ABC** – Approved Budget for the Contract.

**BAC** – Bids and Awards Committee.

**Bid** – A signed offer or proposal to undertake a contract submitted by a bidder in response to and in consonance with the requirements of the bidding documents. Also referred to as *Proposal* and *Tender*. (2016 revised IRR, Section 5[c])

**Bidder** – Refers to a contractor, manufacturer, supplier, distributor and/or consultant who submits a bid in response to the requirements of the Bidding Documents. (2016 revised IRR, Section 5[d])

**Bidding Documents** – The documents issued by the Procuring Entity as the bases for bids, furnishing all information necessary for a prospective bidder to prepare a bid for the Goods, Infrastructure Projects, and/or Consulting Services required by the Procuring Entity. (2016 revised IRR, Section 5[e])

**BIR** – Bureau of Internal Revenue.

**BSP** – Bangko Sentral ng Pilipinas.

**Consulting Services** – Refer to services for Infrastructure Projects and other types of projects or activities of the GOP requiring adequate external technical and professional expertise that are beyond the capability and/or capacity of the GOP to undertake such as, but not limited to: (i) advisory and review services; (ii) pre-investment or feasibility studies; (iii) design; (iv) construction supervision; (v) management and related services; and (vi) other technical services or special studies. (2016 revised IRR, Section 5[i])

**CDA** - Cooperative Development Authority.

**Contract** – Refers to the agreement entered into between the Procuring Entity and the Supplier or Manufacturer or Distributor or Service Provider for procurement of Goods and Services; Contractor for Procurement of Infrastructure Projects; or Consultant or Consulting Firm for Procurement of Consulting Services; as the case may be, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.

**CIF** – Cost Insurance and Freight.

**CIP** – Carriage and Insurance Paid.

**CPI** – Consumer Price Index.

**DDP** – Refers to the quoted price of the Goods, which means “delivered duty paid.”

**DTI** – Department of Trade and Industry.

**EXW** – Ex works.

**FCA** – “Free Carrier” shipping point.

**FOB** – “Free on Board” shipping point.

**Foreign-funded Procurement or Foreign-Assisted Project**– Refers to procurement whose funding source is from a foreign government, foreign or international financing institution as specified in the Treaty or International or Executive Agreement. (2016 revised IRR, Section 5[b]).

**Framework Agreement** – Refers to a written agreement between a procuring entity and a supplier or service provider that identifies the terms and conditions, under which specific purchases, otherwise known as “Call-Offs,” are made for the duration of the agreement. It is in the nature of an option contract between the procuring entity and the bidder(s) granting the procuring entity the option to either place an order for any of the goods or services identified in the Framework Agreement List or not buy at all, within a minimum period of one (1) year to a maximum period of three (3) years. (GPPB Resolution No. 27-2019)

**GFI** – Government Financial Institution.

**GOCC** – Government-owned and/or –controlled corporation.

**Goods** – Refer to all items, supplies, materials and general support services, except Consulting Services and Infrastructure Projects, which may be needed in the transaction of public businesses or in the pursuit of any government undertaking, project or activity, whether in the nature of equipment, furniture, stationery, materials for construction, or personal property of any kind, including non-personal or contractual services such as the repair and maintenance of equipment and furniture, as well as trucking, hauling, janitorial, security, and related or analogous services, as well as procurement of materials and supplies provided by the Procuring Entity for such services. The term “related” or “analogous services” shall include, but is not limited to, lease or purchase of office space, media advertisements, health maintenance services, and other services essential to the operation of the Procuring Entity. (2016 revised IRR, Section 5[r])

**GOP** – Government of the Philippines.

**GPPB** – Government Procurement Policy Board.

**INCOTERMS** – International Commercial Terms.

**Infrastructure Projects** – Include the construction, improvement, rehabilitation, demolition, repair, restoration or maintenance of roads and bridges, railways, airports, seaports, communication facilities, civil works components of information technology projects, irrigation, flood control and drainage, water supply, sanitation, sewerage and solid waste management systems, shore protection, energy/power and electrification facilities, national buildings, school buildings, hospital buildings, and other related construction projects of the government. Also referred to as *civil works or works*. (2016 revised IRR, Section 5[u])

**LGUs** – Local Government Units.

**NFCC** – Net Financial Contracting Capacity.

**NGA** – National Government Agency.

**PhilGEPS** - Philippine Government Electronic Procurement System.

**Procurement Project** – refers to a specific or identified procurement covering goods, infrastructure project or consulting services. A Procurement Project shall be described, detailed, and scheduled in the Project Procurement Management Plan prepared by the agency

which shall be consolidated in the procuring entity's Annual Procurement Plan. (GPPB Circular No. 06-2019 dated 17 July 2019)

**PSA** – Philippine Statistics Authority.

**SEC** – Securities and Exchange Commission.

**SLCC** – Single Largest Completed Contract.

**Supplier** – refers to a citizen, or any corporate body or commercial company duly organized and registered under the laws where it is established, habitually established in business and engaged in the manufacture or sale of the merchandise or performance of the general services covered by his bid. (Item 3.8 of GPPB Resolution No. 13-2019, dated 23 May 2019). Supplier as used in these Bidding Documents may likewise refer to a distributor, manufacturer, contractor, or consultant.

**UN** – United Nations.



*Section I. Invitation to Bid*

Procurement of  
**MEDICAL  
EQUIPMENT**



**Republic of the Philippines  
Provincial Government of Palawan**

***INVITATION TO BID FOR THE Procurement of MEDICAL  
EQUIPMENT***

1. The **Provincial Government of Palawan**, through the **Trust Fund (Fund Code: TF DUE TO NGAS - DOH)** intends to apply the sum of **Four Million Five Hundred Ninety-Eight Thousand Seven Hundred Ninety-Five Pesos Only (PhP 4,598,795.00)** being the ABC to payments under the contract for **Procurement of MEDICAL EQUIPMENT**. Bids received in excess of the ABC shall be automatically rejected at bid opening.
2. The **Provincial Government of Palawan** now invites bids for the above Procurement Project. Delivery of the Goods is required by **on or before fourteen (14) calendar days**. Bidders should have completed, within **two (2) years** from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).
3. Bidding will be conducted through open competitive bidding procedures using a non-discretionary “*pass/fail*” criterion as specified in the 2016 revised Implementing Rules and Regulations (IRR) of Republic Act (RA) No. 9184.

Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA No. 5183.

4. Prospective Bidders may obtain further information **Bids and Awards Committee (BAC) Secretariat, Provincial Government of Palawan** and inspect the Bidding Documents at the address given below during **office hours, Monday to Friday, 8:00 AM to 5:00 PM**.
5. A complete set of Bidding Documents may be acquired by interested Bidders on **September 7, 2021** from the given address and website(s) below and upon payment of the applicable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB, amounting to **Five Thousand Pesos Only (P 5,000.00)**.
6. The **Provincial Government of Palawan** will hold a Pre-Bid Conference<sup>1</sup> on **September 15, 2021 at 9:30 AM** at the **GSO Conference Room, 3/F Alfredo M. Abueg Sr. Building, Capitol Compound, Puerto Princesa City, Palawan**, which shall be open to prospective bidders.
7. Bids must be duly received by the BAC Secretariat through manual submission at the office address indicated below, **on or before 9:00 AM of September 29, 2021, at the Office of the BAC Secretariat, 3/F Alfredo M. Abueg S. Building, Capitol Compound, Puerto Princesa City, Palawan**. Late bids shall not be accepted.

8. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in **ITB** Clause 14.
9. Bid opening shall be on **9:30 AM of September 29, 2021** at **GSO Conference Room, 3/F Alfredo M. Abueg Sr. Building, Capitol Compound, Puerto Princesa City, Palawan**. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.
10. Schedule of Activities:

ACTIVITIES	PLACE	DATE & TIME
1. Issuance of Bid Documents	Office of the BAC Secretariat, 2/F Provincial Budget Office, Capitol Compound	September 7, 2021
2. Pre-bid Conference	GSO Conference Room, 3/F Alfredo M. Abueg Sr. Building, Capitol Compound	September 15, 2021, at 9:30 AM
3. Deadline of Submission of Bids	Office of the BAC Secretariat, 3/F Alfredo M. Abueg Sr. Building, Capitol Compound	On or before 9:00 AM of September 29, 2021
4. Opening of Bids	GSO Conference Room, 3/F Alfredo M. Abueg Sr. Building, Capitol Compound	September 29, 2021, at 9:30 AM
5. Bid Evaluation		

11. The **Provincial Government of Palawan** reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised IRR of RA No. 9184, without thereby incurring any liability to the affected bidder or bidders.
12. For further information, please refer to:

**ARLENE BORLAGDATAN-LIBRES**  
 Head, Bids and Awards Committee (BAC) Secretariat  
 Provincial Legal Office  
 Provincial Capitol Compound, Puerto Princesa City, Palawan, 5300  
 (048) 433-2948

13. You may visit the following websites:

For downloading of Bidding Documents: <https://notices.philgeps.gov.ph> and <http://palawan.gov.ph/bids/>

[Date of Issue September 7, 2021]

**ATTY. JETHRO M. PALAYON**



*BAC - Chairman*



*Section II. Instructions to Bidders*

Procurement of  
**MEDICAL  
EQUIPMENT**

## 1. Scope of Bid

The Procuring Entity, *Provincial Government of Palawan* wishes to receive Bids for **the Procurement of MEDICAL EQUIPMENT** under a Framework Agreement}, with identification number *TF DUE TO NGAS - DOH 21-011*.

The Procurement Project (referred to herein as “Project”) is composed of lot items, the details of which are described in Section VII (Technical Specifications).

## 2. Funding Information

2.1. The GOP through the source of funding as indicated below for **Trust Fund (Fund Code: TF DUE TO NGAS - DOH)** in the amount of **(PhP 4,598,795.00)**.

2.2. The source of funding is:

*[If an early procurement activity, select one and delete others:]*

a. LGUs, the proposed Local Expenditure Program.

*[If not an early procurement activity, select one and delete others:]*

a. LGUs, the Annual or Supplemental Budget, as approved by the Sanggunian.

## 3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

## 4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex “I” of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

## 5. Eligible Bidders

5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.

- 5.2. Foreign ownership limited to those allowed under the rules may participate in this Project.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:

*[Select one, delete the other/s]*

- a. For the procurement of Non-expendable Supplies and Services: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.
- b. For the procurement of Expendable Supplies: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least twenty-five percent (25%) of the ABC.
- c. For procurement where the Procuring Entity has determined, after the conduct of market research, that imposition of either (a) or (b) will likely result to failure of bidding or monopoly that will defeat the purpose of public bidding: the Bidder should comply with the following requirements: *[Select either failure or monopoly of bidding based on market research conducted]*
- i. Completed at least two (2) similar contracts, the aggregate amount of which should be equivalent to at least *fifty percent (50%) in the case of non-expendable supplies and services or twenty-five percent (25%) in the case of expendable supplies* of the ABC for this Project; and
- ii. The largest of these similar contracts must be equivalent to at least half of the percentage of the ABC as required above.

- 5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

## **6. Origin of Goods**

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **ITB** Clause 18.

## **7. Subcontracts**

- 7.1. The Bidder may subcontract portions of the Project to the extent allowed by the Procuring Entity as stated herein, but in no case more than twenty percent (20%) of the Project.

The Procuring Entity has prescribed that:

Subcontracting is not allowed.

- 7.2. *[If Procuring Entity has determined that subcontracting is allowed during the bidding, state:]* The Bidder must submit together with its Bid the documentary requirements of the subcontractor(s) complying with the eligibility criteria

stated in **ITB** Clause 5 in accordance with Section 23.4 of the 2016 revised IRR of RA No. 9184 pursuant to Section 23.1 thereof.

- 7.3. *[If subcontracting is allowed during the contract implementation stage, state:]* The Supplier may identify its subcontractor during the contract implementation stage. Subcontractors identified during the bidding may be changed during the implementation of this Contract. Subcontractors must submit the documentary requirements under Section 23.1 of the 2016 revised IRR of RA No. 9184 and comply with the eligibility criteria specified in **ITB** Clause 5 to the implementing or end-user unit.
- 7.4. Subcontracting of any portion of the Project does not relieve the Supplier of any liability or obligation under the Contract. The Supplier will be responsible for the acts, defaults, and negligence of any subcontractor, its agents, servants, or workmen as fully as if these were the Supplier's own acts, defaults, or negligence, or those of its agents, servants, or workmen.

## **8. Pre-Bid Conference**

The Procuring Entity will hold a pre-bid conference for this Project on the specified date and time and either at its physical address **GSO Conference Room, 3/F Alfredo M. Abueg Sr. Building, Capitol Compound, Puerto Princesa City, Palawan** and/or through videoconferencing/webcasting } as indicated in paragraph 6 of the **IB**.

## **9. Clarification and Amendment of Bidding Documents**

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **IB**, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

## **10. Documents comprising the Bid: Eligibility and Technical Components**

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within *[state relevant period as provided in paragraph 2 of the **IB**]* prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

## **11. Documents comprising the Bid: Financial Component**

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the **IB** shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.
- 11.5. Financial proposals for single or multi-year Framework Agreement shall be submitted before the deadline of submission of bids as prescribed in the **IB**. For multi-year Framework Agreement, evaluation of the financial proposal during this stage is for purposes of determining eligibility and whether or not such financial proposal is within the ABC.

## 12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
  - a. For Goods offered from within the Procuring Entity's country:
    - i. The price of the Goods quoted EXW (ex-works, ex-factory, ex-warehouse, ex-showroom, or off-the-shelf, as applicable);
    - ii. The cost of all customs duties and sales and other taxes already paid or payable;
    - iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
    - iv. The price of other (incidental) services, if any, listed in e.
  - b. For Goods offered from abroad:
    - i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
    - ii. The price of other (incidental) services, if any, as listed in **Section VII (Technical Specifications)**.
- 12.2. For Framework Agreement, the following should also apply in addition to Clause 12.1:

- a. For a single year Framework Agreement, the prices quoted by the Bidder shall be fixed during the Bidder's performance of the contract and not subject to variation or escalation on any account. Price schedules required under Clause 12.1 shall be submitted with the bidding documents.
- b. For a multi-year Framework Agreement, the prices quoted by the Bidder during submission of eligibility documents shall be the ceiling and the price quoted during mini-competition must not exceed the initial price offer. The price quoted during call for mini-competition shall be fixed during the Bidder's performance of that Call-off and not subject to variation or escalation on any account. Price schedules required under Clause 12.1 shall be submitted with the bidding documents.

### **13. Bid and Payment Currencies**

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.
- 13.2. Payment of the contract price shall be made in:
- 13.3.
  - a. Philippine Pesos.

### **14. Bid Security**

- 14.1. The Bidder shall submit a Bid Securing Declaration<sup>2</sup> or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and bid security shall be valid until one hundred twenty (120) calendar days from the date of the opening of bids. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.
- 14.3. In the case of Framework Agreement, other than the grounds for forfeiture under the 2016 revised IRR, the bid security may also be forfeited if the successful bidder fails to sign the Framework Agreement, or fails to furnish the performance security or performance securing declaration. Without prejudice on its forfeiture, bid securities shall be returned only after the posting of performance security or performance securing declaration, as the case may be, by the winning Bidder or compliant Bidders and the signing of the Framework Agreement.

### **15. Sealing and Marking of Bids**

Each Bidder shall submit one copy of the first and second components of its Bid.

---

<sup>2</sup> In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

## **16. Deadline for Submission of Bids**

- 16.1. The Bidders shall submit on the specified date and time and either at its physical address or through online submission as indicated in paragraph 7 of the **IB**.
- 16.2. For multi-year Framework Agreement, the submission of bids shall be for the initial evaluation of their technical and financial eligibility. Thereafter, those declared eligible during the said initial eligibility evaluation and entered into a Framework Agreement with the Procuring Entity shall submit anew their best financial offer at the address and on or before the date and time indicated in the Call for each mini-competition.

## **17. Opening and Preliminary Examination of Bids**

- 17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IB**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

- 17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

## **18. Domestic Preference**

- 18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

## **19. Detailed Evaluation and Comparison of Bids**

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "*passed*," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.
- 19.3. The descriptions of the lots or items shall be indicated in **Section VII (Technical Specifications)**, although the ABCs of these lots or items are



indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.

19.4. The Project shall be awarded as follows:

One Project having several items that shall be awarded as one contract.

19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

## **20. Post-Qualification**

20.1. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, the Bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**.

## **21. Signing of the Contract**

21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

21.2. At the same time as the Procuring Entity notifies the successful Bidder that its bid has been accepted, the Procuring Entity shall send the Framework Agreement Form to the Bidder, which contract has been provided in the Bidding Documents, incorporating therein all agreements between the parties.

21.3. Within ten (10) calendar days from receipt of the Notice to Execute Framework Agreement with the Procuring Entity, the successful Bidder or its duly authorized representative shall formally enter into a Framework Agreement.



*Section III. Bid Data Sheet*

Procurement of  
**MEDICAL  
EQUIPMENT**

# Bid Data Sheet

ITB Clause	
5.3	<p>For this purpose, contracts similar to the Project shall be:</p> <p>a. <i>[provide the definition or description of similar contracts].</i></p> <p>b. completed within <i>[indicate period]</i> prior to the deadline for the submission and receipt of bids.</p>
7.1	<i>Subcontracting is not allowed.</i>
12	The price of the Goods shall be quoted DDP or the applicable International Commercial Terms (INCOTERMS) for this Project.
14.1	<p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <p>a. The amount of not less than two percent (2%) of the ABC, if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or</p> <p>b. The amount of not less than five percent (5%) of the ABC if bid security is in Surety Bond.</p>
15	<i>Bids submitted online or any other electronic means must be accompanied by a proof or receipt that the hard copies of documents were sent on or before the scheduled time and date of Opening of Bids.</i>
19.3	<p><i>[In case the Project will be awarded by lot, list the grouping of lots by specifying the group title, items, and the quantity for every identified lot, and the corresponding ABC for each lot.]</i></p> <p><i>[In case the project will be awarded by item, list each item indicating its quantity and ABC.]</i></p>
20.2	<p><i>Licenses, if applicable:</i></p> <p><i>(a) Sanitary Permit issued by the Local Health Department, for Food / Catering Establishments (for Procurement of Food / Catering Services)</i></p> <p><i>(b) FDA License to Operate as Drug Distributor / Wholesaler (for Procurement of Drugs &amp; Medicines)</i></p> <p><i>(c) Updated PDEA S-4 License (if the procurement includes dangerous or regulated drugs and substances)</i></p> <p><i>(d) FDA License to Operate as Dealer of Medical &amp; Dental Supplies (for Procurement of Medical &amp; Dental Supplies)</i></p> <p><i>(e) FDA License to Operate as Dealer of Medical &amp; Dental Equipment</i></p>

	<p><i>(for Procurement of Medical &amp; Dental Equipment)</i></p> <p><i>(f) SAGSD License to Operate (for the Procurement of Security Services)</i></p> <p><i>(g) Bureau of Animal Industry (BAI) accreditation Certificate (for Procurement of Veterinary / Livestock Supplies)</i></p> <p><i>(h) Bureau of Plant Industry (BPI) accreditation Certificate (for Procurement of Seeding / Agricultural Supplies)</i></p>
21.2	<i>No additional requirement.</i>



*Section IV. General Conditions of Contract*

Procurement of  
**MEDICAL  
EQUIPMENT**

## 1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC)**.

## 2. Advance Payment and Terms of Payment

2.1. Advance payment of the contract amount is provided under Annex “D” of the revised 2016 IRR of RA No. 9184.

2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the **SCC**.

## 3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184. *{[Include if Framework Agreement will be used:]} In the case of Framework Agreement, the Bidder may opt to furnish the performance security or a Performance Securing Declaration as defined under the Guidelines on the Use of Framework Agreement.*

## 4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project *{[Include if Framework Agreement will be used:]} or Framework Agreement* specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the **SCC, Section IV (Technical Specifications)** shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

## 5. Warranty

- 6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.
- 6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

## **6. Liability of the Supplier**

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.



*Section V. Special Conditions of Contract*

Procurement of  
MEDICAL  
EQUIPMENT



## Special Conditions of Contract

GCC Clause	
1	<p><b>Delivery and Documents –</b></p> <p>For purposes of the Contract, “EXW,” “FOB,” “FCA,” “CIF,” “CIP,” “DDP” and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows:</p> <p>The delivery terms applicable to this Contract are delivered <i>[indicate place of destination]</i>. Risk and title will pass from the Supplier to the Procuring Entity upon receipt and final acceptance of the Goods at their final destination.”</p> <p>Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements).</p> <p>For purposes of this Clause the Procuring Entity’s Representative at the Project Site is <i>[indicate name(s)]</i>.</p> <p><b>Incidental Services –</b></p> <p>The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements: <i>Select appropriate requirements and delete the rest.</i></p> <ol style="list-style-type: none"> <li>a. performance or supervision of on-site assembly and/or start-up of the supplied Goods;</li> <li>b. furnishing of tools required for assembly and/or maintenance of the supplied Goods;</li> <li>c. furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;</li> <li>d. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and</li> </ol>
	<ol style="list-style-type: none"> <li>e. training of the Procuring Entity’s personnel, at the Supplier’s plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.</li> <li>f. <i>[Specify additional incidental service requirements, as needed.]</i></li> </ol> <p>The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.</p> <p><b>Spare Parts –</b></p> <p>The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:</p>

*Select appropriate requirements and delete the rest.*

- a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
- b. in the event of termination of production of the spare parts:
  - i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and
  - ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.

The spare parts and other components required are listed in **Section VI (Schedule of Requirements)** and the cost thereof are included in the contract price.

The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for a period of *[indicate here the time period specified. If not used indicate a time period of three times the warranty period]*.

Spare parts or components shall be supplied as promptly as possible, but in any case, within *[insert appropriate time period]* months of placing the order.

**Packaging –**

The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.

The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.

The outer packaging must be clearly marked on at least four (4) sides as follows:

- Name of the Procuring Entity
- Name of the Supplier
- Contract Description
- Final Destination
- Gross weight
- Any special lifting instructions
- Any special handling instructions
- Any relevant HAZCHEM classifications

	<p>A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.</p> <p><b>Transportation –</b></p> <p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.</p> <p>Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.</p>
	<p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.</p> <p>The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.</p> <p><b>Intellectual Property Rights –</b></p> <p>The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.</p>
2.2	<p><i>[If partial payment is allowed, state]</i> “The terms of payment shall be as follows: _____.”</p>
4	<p>The inspections and tests that will be conducted are: <i>[Indicate the applicable inspections and tests]</i></p>



*Section VI. Schedule of Requirements*

Procurement of  
**MEDICAL  
EQUIPMENT**

# Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item No	Description	Qty	Unit	Total	Delivered, Weeks/Months
1	<p><b>Micro-Centrifuge</b> Features:</p> <ul style="list-style-type: none"> <li>- Wide applicable range: It's suitable for all standard PCR plates with skirted, non-skirted, Elisa Plates, 8x0.2ml PCR strip tubes, and 0.2ml PCR tubes</li> <li>- LCD display, parameters can be modified during operation, user-friendly programming, configuration of jog, timing and speed adjustable operation mode, making the experiment more convenient, efficient, quiet and stable</li> <li>Instantaneous centrifugal function, Instantaneous acceleration to set speed, automatic braking, fast stop time.</li> <li>- DC motor is maintenance-free, faster acceleration and deceleration rate can separate more samples within shorter time.</li> <li>- Humanized design.</li> <li>- Two mode -Pressing for operating and setting time for operating-making the experiment more convenient and efficient.</li> </ul> <p>Smart size, simple operation. The cover can be fully opened to 95°, convenient to put and fetch the plates.; Quickly spin down droplets and condensation. Used before and after thermal cycling to improve PCR yield.</p> <p>Technical specifications: Max Speed: 500-3000r/min; Max RCF: 600×g; Max Capacity: 2 × 96 wells PCR plate; Speed Accuracy: ±20r/min; Temperature Range:0-99min/sec; Noise: ≤50dB; Power Supply: AC 100-230V 50/60HZ 3A; Dimension:230×195×190 (mm) ;Weight: 2.6 kg</p>	4	Unit		Delivery must be <b>on or before fourteen (14) calendar days</b> after the receipt of Notice to Proceed.
2	<p><b>Mini-Centrifuge</b> FEATURES:</p> <ul style="list-style-type: none"> <li>- Strong plastic case with microcomputer control</li> <li>- DC brushless motor, LED display of speed and time</li> <li>- With door lock protection, will stopped when opening the lid.</li> </ul> <p>SPECIFICATIONS:</p> <p>Speed range: 4000-12000 rpm (increase 1000pm); Max RCF: 7900*g; Time range: 1-99 min</p> <p>Capacity: 6*2.0ml/1.5ml/0.5ml/0.2ml tubes 2*8* 0.2mlPCR tube</p> <p>Voltage: 85-265V 50/60Hz; Noise: 45db; Weight:1.5kg; Dimension: 200*180*120 mm</p>	4	Unit		

3	<p><b>Vortex Mixer</b>  Features:  - Various speed controls to the maximum of 3200rpm, provided gentle or vigorous vortexing for your mixing needs.  - 2 interchangeable mixing head i.e. snap on the 26mm diameter rubber cup head for conventional single tube mixing, or a 3" diameter platform pad for multiple tubes and one flask mixing.  2-mode operation features "TOUCH" for intermittent operation, or "SWITCH" activated operation at the flip of a switch for continuous mixing.; Stirring intensity can be controlled in "TOUCH" mode by applying the pressure to the mixing pad.; User-friendly design capable of sustaining to prevent mixer from slipping and creeping due to the vibration  Specifications: Dimension:130 x 160 x 180mm; Eccentric:5.6mmOperating Modes: Touch / Switch; Speed (single tube): 80 ~ 3150rpm (<math>\pm 10\%</math>); Standard accessory: Single tube head; Optional accessory:3" head only; Net weight : 5.0kg</p>	4	Unit	
4	<p><b>Set of Single Channel Pipettors</b>  Features  - Lightweight, ergonomic, low force design; Digital display clearly reads volume setting; The pipettes cover volume range of 0.1<math>\mu</math>l to 10ml; Easy to calibrate and maintain with tool supplied.  - Design helps avoid repetitive strain injuries; Calibrated in accordance with ISO8655; Each pipette supplied with individual test certificate; The low part is available for autoclaving.  Technical Data:  Volume Range Increment Test Volume (<math>\mu</math>l) Maximum Permissible Systematic error; (Inaccuracy) Maximum Permissible; Random error (Inaccuracy) % <math>\mu</math>l % <math>\mu</math>l</p> <p>0.1-2.5<math>\mu</math>l 0.05<math>\mu</math>l 2.5 2.50% 0.0625 2.00% 0.05  1.25 3.00% 0.0375 3.00% 0.0375  0.25 12.00% 0.03 6.00% 0.015  0.5-10<math>\mu</math>l 0.1<math>\mu</math>l 10 1.00% 0.1 0.80% 0.08  5 1.50% 0.075 1.50% 0.075; 1 2.50% 0.025 1.50% 0.015</p> <p>2-20<math>\mu</math>l 0.5<math>\mu</math>l 20 0.90% 0.18 0.40% 0.08  10 1.20% 0.12 1.00% 0.1; 2 3.00% 0.06 2.00% 0.04  5-50<math>\mu</math>l 0.5<math>\mu</math>l 50 0.60% 0.3 0.30% 0.15  25 0.90% 0.225 0.60% 0.15; 5 2.00% 0.1 2.00% 0.1  10-100<math>\mu</math>l 1<math>\mu</math>l 100 0.80% 0.8 0.15% 0.15  50 1.00% 0.5 0.40% 0.2; 10 3.00% 0.3 1.50% 0.15</p>	10	Unit	

	<p>20-200µl 1µl 200 0.60% 1.2 0.15% 0.3  100 0.80% 0.8 0.30% 0.3; 20 3.00% 0.6  1.00% 0.2  50-200µl 1µl 200 0.60% 1.2 0.15% 0.3  100 0.80% 0.8 0.30% 0.3; 50 1.00% 0.5  0.40% 0.2  Inclusion: Pipette rack per set of pipettors</p>			
5	<p><b>Microplate spinner, 96 well PCR plate, 2 plate capacity</b>  Features:  - Wide applicable range: It's suitable for all standard PCR plates with skirted, non-skirted, Elisa Plates, 8x0.2ml PCR strip tubes, and 0.2ml PCR tubes  - LCD display, parameters can be modified during operation, user-friendly programming, configuration of jog, timing and speed adjustable operation mode, making the experiment more convenient, efficient, quiet and stable</p> <p>Instantaneous centrifugal function, instantaneous acceleration to set speed, automatic braking, fast stop time.  - DC motor is maintenance-free, faster acceleration and deceleration rate can separate more samples within shorter time.  - Humanized design; Two mode -Pressing for operating and setting time for operating-making the experiment more convenient and efficient.</p> <p>Smart size, simple operation. The cover can be fully opened to 95°, convenient to put and fetch the plates.; Quickly spin down droplets and condensation. Used before and after thermal cycling to improve PCR yield.  Technical specifications:  Max Speed: 500-3000r/min; Max RCF: 600×g;  Max Capacity: 2 × 96 wells PCR plate; Speed Accuracy: ±20r/min  Temperature Range:0-99min/sec; Noise: ≤50dB; Power Supply: AC 100-230V 50/60HZ 3A; Dimension:230×195×190 (mm)  Weight: 2.6 kg</p>	2	Unit	
6	<p><b>Powered Air-Purifying Respirator</b>  Specification:  Double-Sided Anti-Fog; Anti-Impact and Anti-Scratch; Large lens provides a wide field of view for excellent visibility; Head Net design; Fast Wearing and Good Air Tightness; Not easy to produce sliding and gasps; No pressure when wearing; Cartridge will last up to 3 months (depending upon of the Environment).</p>	8	Unit	
7	<p><b>Air Filtered Pass Box</b>  Type: Air Shower Pass Box  External size: 660 x 570 x 1000 mm; Internal size: 500 x 500 x 500 mm; Materials: Stainless</p>	8	Unit	

	steel; Interlock: Electrical; UV Lamp: 8W x 1; HEPA Filter: Efficiency ≥99.999% at 0.3um; Clean air speed: 0.36~0.54m/s; Gross Weight: 80Kg			
8	<b>Medical Laboratory Deep Sink</b> * Length: 50cm, Width: 40cm , Depth: 30cm * Thickness: 0.8mm * Stainless Steel * With Soundproof Padding * FREE drain & hose accessories	6	Unit	
9	<b>Ultra-low Temperature Freezer</b> Technical Specification: Structure: External structure in anti-corrosion and non-toxic hot-galvanized sheet metal, covered with PVC film. Internal structure in stainless steel.; Isolation: Insulation thickness of 100 mm., Obtained by injection of polyurethane foam, ecological, CFC-free, high density for optimal conservation of the product and energy saving.; Door: Self-closing blind door with magnetic seal on all four sides for perfect closure. Opening to the right or, on request, to the left.  <b>Interior design:</b> 6 extractable drawers made of stainless steel, mounted on telescopic anti-tipping guides that allow the total exit of the drawer.; Transparent front in plexiglass.; Temperature: Adjustable temperature range: -10 ° C / -42 ° C. Factory setting: -40 ° C; Refrigerant circuit: Closed circuit hermetic compressor, silent, with thermal protector and finned pack condensing system with helicoidal fan motor fan (suitable for tropical climates). Static evaporator. Ecological refrigerant gas free of CFC.  <b>Defrosting:</b> Defrosting at machine standstill.; Capacity: 600 liters; Voltage:230V / 50 Hz (on request 230V / 60 Hz, 115V / 60 Hz); Absorption:580 Watt; Gas: CFC free; Control panel and alarms:Microprocessor control panel with digital display and easy internal temperature display. Temperature control by PT1000 probe. Audible and visual alarms: high and low temperature (adjustable threshold), open door, condenser, faulty probes, power failure, battery.; Data logger:System for recording data relating to temperature and alarms of the last 30 days that can be downloaded on a USB key.  <b>Standard equipment:</b> 6 drawers in stainless steel; four unidirectional wheels with stabilizing feet; 3x1 separable power cable complete with schuko plug (on request English plug); electric anti-interference filter;protection fuses;	2	Unit	



	<p>switching on and off with a password; electronic control panel with alarm system; graphical weekly temperature recorder, operating with 1.5V battery, equipped with 52 disks and double nib; control unit powered by a 12V buffer battery for temperature display and alarm acquisition in the absence of electricity; contact for remote alarm; USB port; alarm memory.;External dimensions and weight: 780x840x1930 (WxDxH - mm) - kg.173.</p> <p><b>Dimensions and weight with packaging:</b> cardboard box: 910x850x2100 (WxDxH - mm) - kg. 186; wooden case (optional on request):1020x930x2200 (WxDxH - mm) - kg. 243; Equipment compliant with: Low Voltage Directive 2014/35 / EU, Electromagnetic Compatibility Directive 2014/30 / UE.This appliance complies with the essential requirements of Directive 93/42 / EEC and subsequent amendments and additions. Medical device in class IIa; Quality Standard:ISO9001 / 14001 / 13485, CE Marking;</p>			
10	<p><b>Ultra-low Temperature Freezer PCR hood/ Workstation</b> Features: UV sterilization system; HEPA filter efficiency 99.999% at 0.3µm; Interlock function: UV lamp only can be switched on when the front window is closed, ensuring operator safety; UV timer (0-90minutes): When the setting time expires, the UV lamp automatically switches off in preparation for the next experiment</p> <p>Technical Parameters: External Size: 800 x 700 x 1770 mm; Internal Size: 700 x 595 x 550 mm; Work Surface Height:750mm Material: Main Body: Cold-rolled steel with anti-bacteria powder coating; Work Table: 304 stainless steel; Front &amp; Side Windows: 5mm toughened glass , anti-UV , motorized; Max Opening: 300mm; Pre-Filter: Polyester fiber, Washable; HEPA Filter: 99.999% efficiency at 0.3µm.</p> <p>Airflow Velocity: 0.3-0.5m/s, speed adjustable; Noise: ≤65dB; UV Lamp 20W x 1 Emission of 253.7 nanometers, with UV timer Illumination Lamp: LED 8W x 2; Illumination: ≥1000Lux Shelf with IV bar: Stainless steel, 502 x 150 x 50mm</p> <p>Caster: Universal caster with leveling feet; Consumption: 400W; Waterproof Socket: Two,</p>	2	Unit	

	total load ≤500W; Power Supply: AC 220V±10%, 50/60Hz;110V±10%,60Hz; Standard Accessory: UV lamp, Fluorescent lamp, Base stand, Waterproof socket x 2, Shelf with IV bar; Gross Weight: 130 kg			
11	<b>Biological Safety Cabinet Class II A2</b> *Double-wall cabinet structures with negative pressure. *Supply/Exhaust HEPA filter 99.99% efficiency *Built-in UV light timer for 0~99 minutes of hours. *Power Consumption: 220V 50/60Hz UV lamp, electrical outlet Set of 4 Brake Casters Technical Specifications Cabinet Type: 4 feet; Internal Dimensions W*D*H 1220*626*709 mm; Air Flow: 70% recirculated / 30% exhausted; Window Operation Height: 20 cm; Ave. Inflow Velocity: 0.51 m/s; Ave. Down Flow Velocity: 0.30 m/s; Supply/Exhaust HEPA Filter: 99.99%UV Lamp: 254 nm / FL 30W*1; Flourecent Lamp: t5 / 28W*2; Window Glass	2	Unit	
12	<b>Fingertip Pulse Oximeter</b> OLED screen with 6 display modes, with soft silicin pads 4 direstion display, resistance to surrounding light and man made light, streamline design, auto shut off without signa after 8s, value signal includes Perfusion Index (PI) 5 levels adjustable brightness	5	Unit	
<i>x-x-x-x Nothing Follows x-x-x-x</i>				
<b>Approved Budget for the Contract</b>				<b>P 4,598,795.00</b>



*Section VII. Technical Specifications*

Procurement of  
**MEDICAL  
EQUIPMENT**

## Technical Specifications

Item No	Specification	Quantity	Unit	Comply or Not Comply
1	<p><b>Micro-Centrifuge</b>  Features:  - Wide applicable range: It's suitable for all standard PCR plates with skirted, non-skirted, Elisa Plates, 8x0.2ml PCR strip tubes, and 0.2ml PCR tubes  - LCD display, parameters can be modified during operation, user-friendly programming, configuration of jog, timing and speed adjustable operation mode, making the experiment more convenient, efficient, quiet and stable  Instantaneous centrifugal function, Instantaneous acceleration to set speed, automatic braking, fast stop time.  - DC motor is maintenance-free, faster acceleration and deceleration rate can separate more samples within shorter time.  - Humanized design.  - Two mode -Pressing for operating and setting time for operating-making the experiment more convenient and efficient.  Smart size, simple operation. The cover can be fully opened to 95°, convenient to put and fetch the plates.; Quickly spin down droplets and condensation. Used before and after thermal cycling to improve PCR yield.  Technical specifications: Max Speed: 500-3000r/min; Max RCF: 600×g; Max Capacity: 2 × 96 wells PCR plate; Speed Accuracy: ±20r/min; Temperature Range:0-99min/sec; Noise: ≤50dB; Power Supply: AC 100-230V 50/60HZ 3A; Dimension:230×195×190 ( mm) ;Weight: 2.6 kg</p>	4	Unit	
2	<p><b>Mini-Centrifuge</b>  FEATURES:  - Strong plastic case with microcomputer control  - DC brushless motor, LED display of speed and time  - With door lock protection, will stopped when opening the lid.  SPECIFICATIONS:  Speed range: 4000-12000 rpm (increase 1000pm);  Max RCF: 7900*g; Time range: 1-99 min  Capacity: 6*2.0ml/1.5ml/0.5ml/0.2ml tubes 2*8*0.2mlPCR tube  Voltage: 85-265V 50/60Hz; Noise: 45db;  Weight:1.5kg; Dimension: 200*180*120 mm</p>	4	Unit	
3	<p><b>Vortex Mixer</b>  Features:  - Various speed controls to the maximum of 3200rpm, provided gentle or vigorous vortexing for your mixing needs.  - 2 interchangeable mixing head i.e. snap on the 26mm diameter rubber cup head for conventional</p>	4	Unit	

	<p>single tube mixing, or a 3" diameter platform pad for multiple tubes and one flask mixing.</p> <p>2-mode operation features "TOUCH" for intermittent operation, or "SWITCH" activated operation at the flip of a switch for continuous mixing.; Stirring intensity can be controlled in "TOUCH" mode by applying the pressure to the mixing pad.; User-friendly design capable of sustaining to prevent mixer from slipping and creeping due to the vibration</p> <p>Specifications: Dimension:130 x 160 x 180mm; Eccentric:5.6mmOperating Modes: Touch / Switch; Speed (single tube): 80 ~ 3150rpm (<math>\pm 10\%</math>); Standard accessory: Single tube head; Optional accessory:3" head only; Net weight : 5.0kg</p>			
4	<p><b>Set of Single Channel Pipettors</b></p> <p>Features</p> <ul style="list-style-type: none"> <li>- Lightweight, ergonomic, low force design; Digital display clearly reads volume setting; The pipettes cover volume range of 0.1µl to 10ml; Easy to calibrate and maintain with tool supplied.</li> <li>- Design helps avoid repetitive strain injuries; Calibrated in accordance with ISO8655; Each pipette supplied with individual test certificate; The low part is available for autoclaving.</li> </ul> <p>Technical Data:</p> <p>Volume Range Increment Test Volume (µl) Maximum Permissible Systematic error; (Inaccuracy) Maximum Permissible; Random error (Inaccuracy) % µl % µl</p> <p>0.1-2.5µl 0.05µl 2.5 2.50% 0.0625 2.00% 0.05 1.25 3.00% 0.0375 3.00% 0.0375 0.25 12.00% 0.03 6.00% 0.015</p> <p>0.5-10µl 0.1µl 10 1.00% 0.1 0.80% 0.08 5 1.50% 0.075 1.50% 0.075; 1 2.50% 0.025 1.50% 0.015</p> <p>2-20µl 0.5µl 20 0.90% 0.18 0.40% 0.08 10 1.20% 0.12 1.00% 0.1; 2 3.00% 0.06 2.00% 0.04</p> <p>5-50µl 0.5µl 50 0.60% 0.3 0.30% 0.15 25 0.90% 0.225 0.60% 0.15; 5 2.00% 0.1 2.00% 0.1</p> <p>10-100µl 1µl 100 0.80% 0.8 0.15% 0.15 50 1.00% 0.5 0.40% 0.2; 10 3.00% 0.3 1.50% 0.15</p> <p>20-200µl 1µl 200 0.60% 1.2 0.15% 0.3 100 0.80% 0.8 0.30% 0.3; 20 3.00% 0.6 1.00% 0.2</p> <p>50-200µl 1µl 200 0.60% 1.2 0.15% 0.3 100 0.80% 0.8 0.30% 0.3; 50 1.00% 0.5 0.40% 0.2</p> <p>Inclusion: Pipette rack per set of pipettors</p>	10	Unit	
5	<p><b>Microplate spinner, 96 well PCR plate, 2 plate capacity</b></p> <p>Features:</p> <ul style="list-style-type: none"> <li>- Wide applicable range: It's suitable for all standard PCR plates with skirted, non-skirted, Elisa Plates, 8x0.2ml PCR strip tubes, and 0.2ml PCR tubes</li> <li>- LCD display, parameters can be modified during</li> </ul>	2	Unit	

	<p>operation, user-friendly programming, configuration of jog, timing and speed adjustable operation mode, making the experiment more convenient, efficient, quiet and stable</p> <p>Instantaneous centrifugal function, instantaneous acceleration to set speed, automatic braking, fast stop time.</p> <ul style="list-style-type: none"> <li>- DC motor is maintenance-free, faster acceleration and deceleration rate can separate more samples within shorter time.</li> <li>- Humanized design; Two mode -Pressing for operating and setting time for operating-making the experiment more convenient and efficient.</li> </ul> <p>Smart size, simple operation. The cover can be fully opened to 95°, convenient to put and fetch the plates.; Quickly spin down droplets and condensation. Used before and after thermal cycling to improve PCR yield.</p> <p>Technical specifications:  Max Speed: 500-3000r/min; Max RCF: 600×g; Max Capacity: 2 × 96 wells PCR plate; Speed Accuracy: ±20r/min  Temperature Range:0-99min/sec; Noise: ≤50dB;  Power Supply: AC 100-230V 50/60HZ 3A;  Dimension:230×195×190 (mm)  Weight: 2.6 kg</p>			
6	<p><b>Powered Air-Purifying Respirator</b>  Specification:  Double-Sided Anti-Fog; Anti-Impact and Anti-Scratch;  Large lens provides a wide field of view for excellent visibility; Head Net design; Fast Wearing and Good Air Tightness; Not easy to produce sliding and gasps;  No pressure when wearing; Cartridge will last up to 3 months (depending upon of the Environment).</p>	8	Unit	
7	<p><b>Air Filtered Pass Box</b>  Type: Air Shower Pass Box  External size: 660 x 570 x 1000 mm; Internal size: 500 x 500 x 500 mm; Materials: Stainless steel;  Interlock: Electronical; UV Lamp: 8W x 1; HEPA Filter: Efficiency ≥99.999% at 0.3um; Clean air speed: 0.36~0.54m/s; Gross Weight: 80Kg</p>	8	Unit	
8	<p><b>Medical Laboratory Deep Sink</b>  * Length: 50cm, Width: 40cm , Depth: 30cm  * Thickness: 0.8mm  * Stainless Steel  * With Soundproof Padding  * FREE drain &amp; hose accessories</p>	6	Unit	
9	<p><b>Ultra-low Temperature Freezer</b>  Technical Specification:  Structure: External structure in anti-corrosion and non-toxic hot-galvanized sheet metal, covered with PVC film. Internal structure in stainless steel.;  Isolation: Insulation thickness of 100 mm., Obtained by injection of polyurethane foam, ecological, CFC-free, high density for optimal conservation of the</p>	2	Unit	

	<p>product and energy saving.; Door: Self-closing blind door with magnetic seal on all four sides for perfect closure. Opening to the right or, on request, to the left.</p> <p><b>Interior design:</b> 6 extractable drawers made of stainless steel, mounted on telescopic anti-tipping guides that allow the total exit of the drawer.; Transparent front in plexiglass.; Temperature: Adjustable temperature range: -10 ° C / -42 ° C. Factory setting: -40 ° C; Refrigerant circuit: Closed circuit hermetic compressor, silent, with thermal protector and finned pack condensing system with helicoidal fan motor fan (suitable for tropical climates). Static evaporator. Ecological refrigerant gas free of CFC.</p> <p><b>Defrosting:</b> Defrosting at machine standstill.; Capacity: 600 liters; Voltage:230V / 50 Hz (on request 230V / 60 Hz, 115V / 60 Hz); Absorption:580 Watt; Gas: CFC free; Control panel and alarms:Microprocessor control panel with digital display and easy internal temperature display. Temperature control by PT1000 probe. Audible and visual alarms: high and low temperature (adjustable threshold), open door, condenser, faulty probes, power failure, battery.; Data logger:System for recording data relating to temperature and alarms of the last 30 days that can be downloaded on a USB key.</p> <p><b>Standard equipment:</b> 6 drawers in stainless steel; four unidirectional wheels with stabilizing feet; 3x1 separable power cable complete with schuko plug (on request English plug); electric anti-interference filter;protection fuses; switching on and off with a password; electronic control panel with alarm system; graphical weekly temperature recorder, operating with 1.5V battery, equipped with 52 disks and double nib; control unit powered by a 12V buffer battery for temperature display and alarm acquisition in the absence of electricity; contact for remote alarm; USB port; alarm memory.;External dimensions and weight: 780x840x1930 (WxDxH - mm) - kg.173.</p> <p><b>Dimensions and weight with packaging:</b> cardboard box: 910x850x2100 (WxDxH - mm) - kg. 186; wooden case (optional on request):1020x930x2200 (WxDxH - mm) - kg. 243; Equipment compliant with: Low Voltage Directive 2014/35 / EU, Electromagnetic Compatibility Directive 2014/30 / UE.This appliance complies with the essential requirements of Directive 93/42 / EEC and subsequent amendments and</p>			
10	<p><b>Ultra-low Temperature Freezer PCR hood/ Workstation</b></p> <p>Features: UV sterilization system; HEPA filter efficiency 99.999% at 0.3µm; Interlock function: UV lamp only</p>	2	Unit	

	<p>can be switched on when the front window is closed, ensuring operator safety; UV timer (0-90minutes): When the setting time expires, the UV lamp automatically switches off in preparation for the next experiment</p> <p>Technical Parameters:  External Size: 800 x 700 x 1770 mm; Internal Size: 700 x 595 x 550 mm; Work Surface Height:750mm  Material: Main Body: Cold-rolled steel with anti-bacteria powder coating; Work Table: 304 stainless steel; Front &amp; Side Windows: 5mm toughened glass , anti-UV , motorized; Max Opening: 300mm; Pre-Filter: Polyester fiber, Washable; HEPA Filter: 99.999% efficiency at 0.3µm.</p> <p>Airflow Velocity: 0.3-0.5m/s, speed adjustable; Noise: ≤65dB; UV Lamp 20W x 1 Emission of 253.7 nanometers, with UV timer  Illumination Lamp: LED 8W x 2; Illumination: ≥1000Lux  Shelf with IV bar: Stainless steel, 502 x 150 x 50mm</p> <p>Caster: Universal caster with leveling feet;  Consumption: 400W; Waterproof Socket: Two, total load ≤500W; Power Supply: AC 220V±10%, 50/60Hz;110V±10%,60Hz; Standard Accessory: UV lamp, Fluorescent lamp, Base stand, Waterproof socket x 2, Shelf with IV bar; Gross Weight: 130 kg</p>			
11	<p><b>Biological Safety Cabinet Class II A2</b>  *Double-wall cabinet structures with negative pressure.  *Supply/Exhaust HEPA filter 99.99% efficiency  *Built-in UV light timer for 0~99 minutes of hours.  *Power Consumption: 220V 50/60Hz  UV lamp, electrical outlet  Set of 4 Brake Casters  Technical Specifications  Cabinet Type: 4 feet; Internal Dimensions W*D*H 1220*626*709 mm; Air Flow: 70% recirculated / 30% exhausted; Window Operation Height: 20 cm; Ave. Inflow Velocity: 0.51 m/s; Ave. Down Flow Velocity: 0.30 m/s; Supply/Exhaust HEPA Filter: 99.99%UV Lamp: 254 nm / FL 30W*1; Flourecent Lamp: t5 / 28W*2; Window Glass</p>	2	Unit	
12	<p><b>Fingertip Pulse Oximeter</b>  OLED screen with 6 display modes, with soft silicin pads 4 direstion display, resistance to surrounding light and man made light, streamline design, auto shut off without signa after 8s, value signal includes Perfusion Index (PI) 5 levels adjustable brightness</p>	5	Unit	
<i>x-x-x-x Nothing Follows x-x-x-x</i>				





*Section VIII. Checklist of Technical and  
Financial Documents/ Bidding Forms*

Procurement of  
**MEDICAL  
EQUIPMENT**



Republic of the Philippines  
 Provincial Government of Palawan  
**BIDS AND AWARDS COMMITTEE**

**CHECKLIST OF REQUIREMENTS FOR BIDDING OF GOODS AND SERVICES**  
 (Updated October 14, 2020)

**Project Ref No.** \_\_\_\_\_ **ABC :** \_\_\_\_\_  
**Project Description:** Procurement of \_\_\_\_\_  
**Requesting Office:** \_\_\_\_\_  
**Date :** \_\_\_\_\_, **Time :** \_\_\_\_\_  
**Place :** BAC Conference Room, 3rd Floor, Alfredo Abueg Sr. Building, Capitol Compound  
**Name of Bidder:** \_\_\_\_\_

**Notes / Instructions:**

1. Documents must be arranged according to this checklist
2. Documents must be properly labeled for ease of identification.

Requirements		Findings	Remarks
<b>FIRST ENVELOPE (Eligibility and Technical Requirements)</b>			
<b>I.</b>	<b>TECHNICAL COMPONENT ENVELOPE</b>		
	<b>Class A Documents-Legal</b>		
A	Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages)		
A.1	<b>OR</b> Registration certificate from Securities and Exchange Commission (SEC), Department of Trade and Industry (DTI) for sole proprietorship or Cooperative Development Authority (CDA) for cooperatives or its equivalent document		
	<b>AND</b> Mayor's or Business Permit issued by the city or municipality where the principal place of business of the prospective bidder is located, or the equivalent document for Exclusive Economic Zones or Areas;		
	<b>AND</b> Tax clearance per EO No. 398, s. 2006 as finally reviewed and approved by the Bureau of Internal Revenue (BIR)		
	<b>Class A Documents-Technical</b>		
B	Statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or nor similar in nature and complexity to the contract to be bid with attachments to support the statement		
C	Statement of the Bidder's Single Largest Completed Contract (SLCC) similar to the contract to be bid, except under conditions provided for in Sections 23.4.1.3 and 23.4.2.4 of the 2016 revised IRR of RA 1984, within the relevant period as provided in the Bidding documents with attachments to support the existence of the transaction		
D	Original copy of Notarized Bid Securing Declaration		
D.1	<b>OR</b> Original copy of Bid Security. If in the form of Security Bond, submit also a certification issued by the Insurance Commission		
E	Conformity with the Technical Specifications, which may include production/ delivery schedule, manpower requirements,		

	and/or after sales/ parts, if applicable		
F	Original duly signed Omnibus Sworn Statement		
G	<b>If applicable.</b> Original Notarized Secretary's Certificate in case of a corporation, partnership or cooperative or Original Special Power of Attorney of all members of the joint venture giving full power and authority to its officer to sign the OSS and do acts to represent the Bidder		
	<b>Financial Documents</b>		
H	Audited Financial Statements, showing among others, the total and current assets and liabilities, stamped "received" by the BIR or its duly accredited and authorized institutions, for the preceding calendar year which should not be earlier than two (2) years from the date of bid submission		
I	Computation of NFCC		
I.1	<b>OR</b> a Committed Line of Credit from a Universal or Commercial Bank in lieu of NFCC computation		
	<b>Class B Documents</b>		
J	A duly signed Joint Venture Agreement (JVA) in case the joint venture is already in existence		
J.1	<b>OR</b> duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful		
K	Other documentary requirement under RA 9184 as applicable		
	For foreign bidders, Certification from the relevant government office of their country stating the Filipinos are allowed to participate in government procurement activities for the same item or product		
	Certification from the DTI if the Bidder claims preference as a Domestic Bidder or Domestic Entity		
<b>II</b>	<b>Post Qualification Requirement : (These requirement will not be check during bid opening but the bidder is encourage to include them in the first envelope to expedite post-qualification in case the bid is the LCB)</b>		
A	Tax Clearance (issued by the Provincial Treasurer) or Certificate of no Landholding (issued by the Provincial Assessor		
B	BIR Certificate of Registration (BIR Form No. 2303)		
C	Philippine Bidding Documents acquired by the Bidder		
D	Latest Income Tax Return		
E	Other Documentary Requirements (Permits & Licenses, Others)		
	i. Sanitary Permit issued by the Local Health Department, for Food / Catering Establishments (for procurement of food / catering services)		
	ii. FDA License to Operate as Drug Distributor / Wholesaler (for procurement of Drugs and Medicines)		
	iii. Updated PDEA S-4 License (if the procurement includes dangerous or regulated drugs and substances)		
	iv. FDA License to Operate as Dealer of Medical & Dental Supplies (for procurement of Medical & Dental Supplies)		
	v. FDA License to Operate as Dealer of Medical & Dental Equipment (for procurement of Medical & Dental Equipment)		
	vi. SAGSD License to Operate (for the Procurement of Security Services)		
	vii. Bureau of Animal Industry (BAI) accreditation		

	Certificate (for procurement of veterinary / livestock supplies)		
	viii. Bureau of Plant Industry (BPI) accreditation Certificate (for procurement of seeding / agricultural supplies)		

<b>SECOND ENVELOPE (Financial Component)</b>			
A	Financial Bid Form		
B	Bill of Quantities		

**P – PASSED**

**F – FAILED**

Prepared by:

---

**For Goods Offered From Within the Philippines**

Name of Bidder \_\_\_\_\_ Invitation to Bid Number \_\_\_\_\_. Page \_ of \_\_\_\_\_.

1	2	3	4		5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit of Issue	Unit price EXW per item	Transportation and Insurance and all other costs incidental to delivery, per item	Sales and other taxes payable if Contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)
1	<p><b>Micro-Centrifuge</b>  Features:  - Wide applicable range: It's suitable for all standard PCR plates with skirted, non-skirted, Elisa Plates, 8x0.2ml PCR strip tubes, and 0.2ml PCR tubes  - LCD display, parameters can be modified during operation, user-friendly programming, configuration of jog, timing and speed adjustable operation mode, making the experiment more convenient, efficient, quiet and stable  Instantaneous centrifugal function, Instantaneous acceleration to set speed, automatic braking, fast stop time.  - DC motor is maintenance-free, faster acceleration and deceleration rate can separate more samples within shorter time.  - Humanized design.  - Two mode -Pressing for operating and setting time for operating-making the experiment more convenient and efficient.  Smart size, simple operation. The cover can be fully opened to 95°, convenient to put and fetch the plates.; Quickly spin down droplets and condensation. Used before and after thermal cycling to improve PCR yield.  Technical specifications: Max Speed: 500-3000r/min; Max RCF: 600×g; Max Capacity: 2 × 96 wells PCR plate; Speed Accuracy: ±20r/min; Temperature Range:0-99min/sec; Noise: ≤50dB; Power Supply: AC 100-230V 50/60HZ 3A;</p>		4	Unit						

	Dimension:230×195×190 (mm) ;Weight: 2.6 kg									
2	<p><b>Mini-Centrifuge</b>  <b>FEATURES:</b>  - Strong plastic case with microcomputer control  - DC brushless motor, LED display of speed and time  - With door lock protection, will stopped when opening the lid.</p> <p><b>SPECIFICATIONS:</b>  Speed range: 4000-12000 rpm (increase 1000rpm); Max RCF: 7900*g; Time range: 1-99 min  Capacity: 6*2.0ml/1.5ml/0.5ml/0.2ml tubes 2*8* 0.2mlPCR tube  Voltage: 85-265V 50/60Hz; Noise: 45db; Weight:1.5kg; Dimension: 200*180*120 mm</p>		4	Unit						
3	<p><b>Vortex Mixer</b>  <b>Features:</b>  - Various speed controls to the maximum of 3200rpm, provided gentle or vigorous vortexing for your mixing needs.  - 2 interchangeable mixing head i.e. snap on the 26mm diameter rubber cup head for conventional single tube mixing, or a 3" diameter platform pad for multiple tubes and one flask mixing.  2-mode operation features "TOUCH" for intermittent operation, or "SWITCH" activated operation at the flip of a switch for continuous mixing.; Stirring intensity can be controlled in "TOUCH" mode by applying the pressure to the mixing pad.; User-friendly design capable of sustaining to prevent mixer from slipping and creeping due to the vibration</p> <p><b>Specifications:</b> Dimension:130 x 160 x 180mm; Eccentric:5.6mmOperating Modes: Touch / Switch; Speed (single tube): 80 ~ 3150rpm (±10%); Standard accessory: Single tube head; Optional accessory:3" head only; Net</p>		4	Unit						

	weight : 5.0kg									
4	<p><b>Set of Single Channel Pipettors</b></p> <p>Features</p> <ul style="list-style-type: none"> <li>- Lightweight, ergonomic, low force design; Digital display clearly reads volume setting; The pipettes cover volume range of 0.1µl to 10ml; Easy to calibrate and maintain with tool supplied.</li> <li>- Design helps avoid repetitive strain injuries; Calibrated in accordance with ISO8655; Each pipette supplied with individual test certificate; The low part is available for autoclaving.</li> </ul> <p>Technical Data:</p> <p>Volume Range Increment Test Volume (µl) Maximum Permissible Systematic error; (Inaccuracy) Maximum Permissible; Random error (Inaccuracy) % µl % µl</p> <p>0.1-2.5µl 0.05µl 2.5 2.50% 0.0625 2.00% 0.05  1.25 3.00% 0.0375 3.00% 0.0375  0.25 12.00% 0.03 6.00% 0.015</p> <p>0.5-10µl 0.1µl 10 1.00% 0.1 0.80% 0.08  5 1.50% 0.075 1.50% 0.075; 1 2.50% 0.025 1.50% 0.015</p> <p>2-20µl 0.5µl 20 0.90% 0.18 0.40% 0.08  10 1.20% 0.12 1.00% 0.1; 2 3.00% 0.06 2.00% 0.04</p> <p>5-50µl 0.5µl 50 0.60% 0.3 0.30% 0.15  25 0.90% 0.225 0.60% 0.15; 5 2.00% 0.1 2.00% 0.1</p> <p>10-100µl 1µl 100 0.80% 0.8 0.15% 0.15  50 1.00% 0.5 0.40% 0.2; 10 3.00% 0.3 1.50% 0.15</p> <p>20-200µl 1µl 200 0.60% 1.2 0.15% 0.3  100 0.80% 0.8 0.30% 0.3; 20 3.00% 0.6 1.00% 0.2</p> <p>50-200µl 1µl 200 0.60% 1.2 0.15% 0.3  100 0.80% 0.8 0.30% 0.3; 50 1.00% 0.5 0.40% 0.2</p>	10	Unit							

	Inclusion: Pipette rack per set of pipettors									
5	<p><b>Microplate spinner, 96 well PCR plate, 2 plate capacity</b>  <b>Features:</b>  - Wide applicable range: It's suitable for all standard PCR plates with skirted, non-skirted, Elisa Plates, 8x0.2ml PCR strip tubes, and 0.2ml PCR tubes  - LCD display, parameters can be modified during operation, user-friendly programming, configuration of jog, timing and speed adjustable operation mode, making the experiment more convenient, efficient, quiet and stable</p> <p>Instantaneous centrifugal function, instantaneous acceleration to set speed, automatic braking, fast stop time.  - DC motor is maintenance-free, faster acceleration and deceleration rate can separate more samples within shorter time.</p> - Humanized design; Two mode -Pressing for operating and setting time for operating-making the experiment more convenient and efficient. <p>Smart size, simple operation. The cover can be fully opened to 95°, convenient to put and fetch the plates.; Quickly spin down droplets and condensation. Used before and after thermal cycling to improve PCR yield.</p> <p>Technical specifications:  Max Speed: 500-3000r/min; Max RCF: 600×g; Max Capacity: 2 × 96 wells PCR plate; Speed Accuracy: ±20r/min  Temperature Range:0-99min/sec; Noise: ≤50dB; Power Supply: AC 100-230V 50/60HZ 3A;  Dimension:230×195×190 (mm)  Weight: 2.6 kg</p>		2	Unit						
6	<b>Powered Air-Purifying Respirator</b>		8	Unit						



	<p>Specification:  Double-Sided Anti-Fog; Anti-Impact and Anti-Scratch; Large lens provides a wide field of view for excellent visibility; Head Net design; Fast Wearing and Good Air Tightness; Not easy to produce sliding and gasps; No pressure when wearing; Cartridge will last up to 3 months (depending upon of the Environment).</p>									
7	<p><b>Air Filtered Pass Box</b>  Type: Air Shower Pass Box  External size: 660 x 570 x 1000 mm; Internal size: 500 x 500 x 500 mm; Materials: Stainless steel; Interlock: Electronical; UV Lamp: 8W x 1; HEPA Filter: Efficiency ≥99.999% at 0.3um; Clean air speed: 0.36~0.54m/s; Gross Weight: 80Kg</p>		8	Unit						
8	<p><b>Medical Laboratory Deep Sink</b>  * Length: 50cm, Width: 40cm , Depth: 30cm  * Thickness: 0.8mm  * Stainless Steel  * With Soundproof Padding  * FREE drain &amp; hose accessories</p>		6	Unit						
9	<p><b>Ultra-low Temperature Freezer</b>  Technical Specification:  Structure: External structure in anti-corrosion and non-toxic hot-galvanized sheet metal, covered with PVC film. Internal structure in stainless steel.; Isolation: Insulation thickness of 100 mm., Obtained by injection of polyurethane foam, ecological, CFC-free, high density for optimal conservation of the product and energy saving.; Door: Self-closing blind door with magnetic seal on all four sides for perfect closure. Opening to the right or, on request, to the left.</p> <p><b>Interior design:</b> 6 extractable drawers made of stainless steel, mounted on telescopic anti-tipping guides that allow</p>		2	Unit						

<p>the total exit of the drawer.; Transparent front in plexiglass.;Temperature: Adjustable temperature range: -10 ° C / -42 ° C. Factory setting: -40 ° C; Refrigerant circuit: Closed circuit hermetic compressor, silent, with thermal protector and finned pack condensing system with helicoidal fan motor fan (suitable for tropical climates). Static evaporator. Ecological refrigerant gas free of CFC.</p> <p><b>Defrosting:</b> Defrosting at machine standstill.; Capacity: 600 liters; Voltage:230V / 50 Hz (on request 230V / 60 Hz, 115V / 60 Hz); Absorption:580 Watt; Gas: CFC free; Control panel and alarms:Microprocessor control panel with digital display and easy internal temperature display. Temperature control by PT1000 probe. Audible and visual alarms: high and low temperature (adjustable threshold), open door, condenser, faulty probes, power failure, battery.; Data logger:System for recording data relating to temperature and alarms of the last 30 days that can be downloaded on a USB key.</p> <p><b>Standard equipment:</b> 6 drawers in stainless steel; four unidirectional wheels with stabilizing feet; 3x1 separable power cable complete with schuko plug (on request English plug); electric anti-interference filter;protection fuses; switching on and off with a password; electronic control panel with alarm system; graphical weekly temperature recorder, operating with 1.5V battery, equipped with 52 disks and double nib; control unit powered by a 12V buffer battery for temperature display and alarm acquisition in the absence of electricity; contact for remote alarm; USB port; alarm memory.;External dimensions and weight: 780x840x1930 (WxDxH - mm) - kg.173.</p> <p><b>Dimensions and weight with packaging:</b> cardboard box:</p>									
--	--	--	--	--	--	--	--	--	--

	910x850x2100 (WxDxH - mm) - kg. 186; wooden case (optional on request):1020x930x2200 (WxDxH - mm) - kg. 243; Equipment compliant with: Low Voltage Directive 2014/35 / EU, Electromagnetic Compatibility Directive 2014/30 / UE.This appliance complies with the essential requirements of Directive 93/42 / EEC and subsequent amendments and additions. Medical device in class IIa; Quality Standard:ISO9001 / 14001 / 13485, CE Marking;								
10	<p><b>Ultra-low Temperature Freezer PCR hood/ Workstation</b></p> <p>Features: UV sterilization system; HEPA filter efficiency 99.999% at 0.3µm; Interlock function: UV lamp only can be switched on when the front window is closed, ensuring operator safety; UV timer (0-90minutes): When the setting time expires, the UV lamp automatically switches off in preparation for the next experiment</p> <p>Technical Parameters: External Size: 800 x 700 x 1770 mm; Internal Size: 700 x 595 x 550 mm; Work Surface Height:750mm Material: Main Body: Cold-rolled steel with anti-bacteria powder coating; Work Table: 304 stainless steel; Front &amp; Side Windows: 5mm toughened glass , anti-UV , motorized; Max Opening: 300mm; Pre-Filter: Polyester fiber, Washable; HEPA Filter: 99.999% efficiency at 0.3µm.</p> <p>Airflow Velocity: 0.3-0.5m/s, speed adjustable; Noise: ≤65dB; UV Lamp 20W x 1 Emission of 253.7 nanometers, with UV timer Illumination Lamp: LED 8W x 2; Illumination: ≥1000Lux Shelf with IV bar: Stainless steel, 502 x 150 x 50mm</p>		2	Unit					

	Caster: Universal caster with leveling feet; Consumption: 400W; Waterproof Socket: Two, total load ≤500W; Power Supply: AC 220V±10%, 50/60Hz; 110V±10%, 60Hz; Standard Accessory: UV lamp, Fluorescent lamp, Base stand, Waterproof socket x 2, Shelf with IV bar; Gross Weight: 130 kg								
11	<b>Biological Safety Cabinet Class II A2</b> *Double-wall cabinet structures with negative pressure. *Supply/Exhaust HEPA filter 99.99% efficiency *Built-in UV light timer for 0~99 minutes of hours. *Power Consumption: 220V 50/60Hz UV lamp, electrical outlet Set of 4 Brake Casters Technical Specifications Cabinet Type: 4 feet; Internal Dimensions W*D*H 1220*626*709 mm; Air Flow: 70% recirculated / 30% exhausted; Window Operation Height: 20 cm; Ave. Inflow Velocity: 0.51 m/s; Ave. Down Flow Velocity: 0.30 m/s; Supply/Exhaust HEPA Filter: 99.99% UV Lamp: 254 nm / FL 30W*1; Fluorescent Lamp: t5 / 28W*2; Window Glass	2	Unit						
12	<b>Fingertip Pulse Oximeter</b> OLED screen with 6 display modes, with soft silicon pads 4 direction display, resistance to surrounding light and man made light, streamline design, auto shut off without signal after 8s, value signal includes Perfusion Index (PI) 5 levels adjustable brightness	5	Unit						
<i>x-x-x-x Nothing Follows x-x-x-x</i>									

\_\_\_\_\_  
Name & Signature of Bidder

\_\_\_\_\_  
Company

**Bid Form for the Procurement of Goods**  
*[shall be submitted with the Bid]*

---

**BID FORM**

Date : \_\_\_\_\_  
 Project Identification No. : \_\_\_\_\_

To: *[name and address of Procuring Entity]*

Having examined the Philippine Bidding Documents (PBDs) including the Supplemental or Bid Bulletin Numbers *[insert numbers]*, the receipt of which is hereby duly acknowledged, we, the undersigned, offer to *[supply/deliver/perform]* *[description of the Goods]* in conformity with the said PBDs for the sum of *[total Bid amount in words and figures]* or the total calculated bid price, as evaluated and corrected for computational errors, and other bid modifications in accordance with the Price Schedules attached herewith and made part of this Bid. The total bid price includes the cost of all taxes, such as, but not limited to: *[specify the applicable taxes, e.g. (i) value added tax (VAT), (ii) income tax, (iii) local taxes, and (iv) other fiscal levies and duties]*, which are itemized herein or in the Price Schedules,

If our Bid is accepted, we undertake:

- a. to deliver the goods in accordance with the delivery schedule specified in the Schedule of Requirements of the Philippine Bidding Documents (PBDs);
- b. to provide a performance security in the form, amounts, and within the times prescribed in the PBDs;
- c. to abide by the Bid Validity Period specified in the PBDs and it shall remain binding upon us at any time before the expiration of that period.

*[Insert this paragraph if Foreign-Assisted Project with the Development Partner:*

Commissions or gratuities, if any, paid or to be paid by us to agents relating to this Bid, and to contract execution if we are awarded the contract, are listed below:

Name and address of agent	Amount	Purpose of Commission or gratuity
_____	_____	_____
_____	_____	_____

(if none, state “None”) ]

Until a formal Contract is prepared and executed, this Bid, together with your written acceptance thereof and your Notice of Award, shall be binding upon us.

We understand that you are not bound to accept the Lowest Calculated Bid or any Bid you may receive.

We certify/confirm that we comply with the eligibility requirements pursuant to the PBDs.

The undersigned is authorized to submit the bid on behalf of *[name of the bidder]* as evidenced by the attached *[state the written authority]*.

We acknowledge that failure to sign each and every page of this Bid Form, including the attached Schedule of Prices, shall be a ground for the rejection of our bid.

Name: \_\_\_\_\_

Legal capacity: \_\_\_\_\_

Signature: \_\_\_\_\_

Duly authorized to sign the Bid for and behalf of: \_\_\_\_\_

Date: \_\_\_\_\_

## **Bid Securing Declaration Form**

*[shall be submitted with the Bid if bidder opts to provide this form of bid security]*

---

REPUBLIC OF THE PHILIPPINES)  
CITY OF \_\_\_\_\_) S.S.

### **BID SECURING DECLARATION** **Project Identification No.: *[Insert number]***

To: *[Insert name and address of the Procuring Entity]*

I/We, the undersigned, declare that:

1. I/We understand that, according to your conditions, bids must be supported by a Bid Security, which may be in the form of a Bid Securing Declaration.
2. I/We accept that: (a) I/we will be automatically disqualified from bidding for any procurement contract with any procuring entity for a period of two (2) years upon receipt of your Blacklisting Order; and, (b) I/we will pay the applicable fine provided under Section 6 of the Guidelines on the Use of Bid Securing Declaration, within fifteen (15) days from receipt of the written demand by the procuring entity for the commission of acts resulting to the enforcement of the bid securing declaration under Sections 23.1(b), 34.2, 40.1 and 69.1, except 69.1(f), of the IRR of RA No. 9184; without prejudice to other legal action the government may undertake.
3. I/We understand that this Bid Securing Declaration shall cease to be valid on the following circumstances:
  - a. Upon expiration of the bid validity period, or any extension thereof pursuant to your request;
  - b. I am/we are declared ineligible or post-disqualified upon receipt of your notice to such effect, and (i) I/we failed to timely file a request for reconsideration or (ii) I/we filed a waiver to avail of said right; and
  - c. I am/we are declared the bidder with the Lowest Calculated Responsive Bid, and I/we have furnished the performance security and signed the Contract.

IN WITNESS WHEREOF, I/We have hereunto set my/our hand/s this \_\_\_\_ day of *[month]* *[year]* at *[place of execution]*.

*[Insert NAME OF BIDDER OR ITS AUTHORIZED REPRESENTATIVE]*

*[Insert signatory's legal capacity]*

Affiant

### **[Jurat]**

*[Format shall be based on the latest Rules on Notarial Practice]*

## Contract Agreement Form for the Procurement of Goods (Revised)

[Not required to be submitted with the Bid, but it shall be submitted within ten (10) days after receiving the Notice of Award]

---

### CONTRACT AGREEMENT

THIS AGREEMENT made the \_\_\_\_\_ day of \_\_\_\_\_ 20\_\_\_\_ between [name of PROCURING ENTITY] of the Philippines (hereinafter called “the Entity”) of the one part and [name of Supplier] of [city and country of Supplier] (hereinafter called “the Supplier”) of the other part;

WHEREAS, the Entity invited Bids for certain goods and ancillary services, particularly [brief description of goods and services] and has accepted a Bid by the Supplier for the supply of those goods and services in the sum of [*contract price in words and figures in specified currency*] (hereinafter called “the Contract Price”).

NOW THIS AGREEMENT WITNESSETH AS FOLLOWS:

1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Conditions of Contract referred to.
2. The following documents as required by the 2016 revised Implementing Rules and Regulations of Republic Act No. 9184 shall be deemed to form and be read and construed as integral part of this Agreement, *viz.*:
  - i. Philippine Bidding Documents (PBDs);
    - i. Schedule of Requirements;
    - ii. Technical Specifications;
    - iii. General and Special Conditions of Contract; and
    - iv. Supplemental or Bid Bulletins, if any
  - ii. Winning bidder’s bid, including the Eligibility requirements, Technical and Financial Proposals, and all other documents or statements submitted;  
  
Bid form, including all the documents/statements contained in the Bidder’s bidding envelopes, as annexes, and all other documents submitted (*e.g.*, Bidder’s response to request for clarifications on the bid), including corrections to the bid, if any, resulting from the Procuring Entity’s bid evaluation;
  - iii. Performance Security;
  - iv. Notice of Award of Contract; and the Bidder’s conforme thereto; and
  - v. Other contract documents that may be required by existing laws and/or the Procuring Entity concerned in the PBDs. **Winning bidder agrees that additional contract documents or information prescribed by the GPPB that are subsequently required for submission after the contract execution, such as the Notice to Proceed, Variation Orders, and Warranty Security, shall likewise form part of the Contract.**
3. In consideration for the sum of [*total contract price in words and figures*] or such other sums as may be ascertained, [*Named of the bidder*] agrees to [*state the object of the contract*] in accordance with his/her/its Bid.
4. The [*Name of the procuring entity*] agrees to pay the above-mentioned sum in accordance with the terms of the Bidding.



IN WITNESS whereof the parties hereto have caused this Agreement to be executed in accordance with the laws of the Republic of the Philippines on the day and year first above written.

*[Insert Name and Signature]*

*[Insert Name and Signature]*

*[Insert Signatory's Legal Capacity]*

*[Insert Signatory's Legal Capacity]*

*for:*

*for:*

*[Insert Procuring Entity]*

*[Insert Name of Supplier]*

**Acknowledgment**

*[Format shall be based on the latest Rules on Notarial Practice]*

## Omnibus Sworn Statement (Revised)

*[shall be submitted with the Bid]*

---

REPUBLIC OF THE PHILIPPINES )  
CITY/MUNICIPALITY OF \_\_\_\_\_ ) S.S.

### AFFIDAVIT

I, [Name of Affiant], of legal age, [Civil Status], [Nationality], and residing at [Address of Affiant], after having been duly sworn in accordance with law, do hereby depose and state that:

1. *[Select one, delete the other:]*

*[If a sole proprietorship:]* I am the sole proprietor or authorized representative of [Name of Bidder] with office address at [address of Bidder];

*[If a partnership, corporation, cooperative, or joint venture:]* I am the duly authorized and designated representative of [Name of Bidder] with office address at [address of Bidder];

2. *[Select one, delete the other:]*

*[If a sole proprietorship:]* As the owner and sole proprietor, or authorized representative of [Name of Bidder], I have full power and authority to do, execute and perform any and all acts necessary to participate, submit the bid, and to sign and execute the ensuing contract for [Name of the Project] of the [Name of the Procuring Entity], as shown in the attached duly notarized Special Power of Attorney;

*[If a partnership, corporation, cooperative, or joint venture:]* I am granted full power and authority to do, execute and perform any and all acts necessary to participate, submit the bid, and to sign and execute the ensuing contract for [Name of the Project] of the [Name of the Procuring Entity], as shown in the attached [state title of attached document showing proof of authorization (e.g., duly notarized Secretary's Certificate, Board/Partnership Resolution, or Special Power of Attorney, whichever is applicable)];

3. [Name of Bidder] is not "blacklisted" or barred from bidding by the Government of the Philippines or any of its agencies, offices, corporations, or Local Government Units, foreign government/foreign or international financing institution whose blacklisting rules have been recognized by the Government Procurement Policy Board, **by itself or by relation, membership, association, affiliation, or controlling interest with another blacklisted person or entity as defined and provided for in the Uniform Guidelines on Blacklisting;**

4. Each of the documents submitted in satisfaction of the bidding requirements is an authentic copy of the original, complete, and all statements and information provided therein are true and correct;

5. [Name of Bidder] is authorizing the Head of the Procuring Entity or its duly authorized representative(s) to verify all the documents submitted;

6. *[Select one, delete the rest:]*

*[If a sole proprietorship:]* The owner or sole proprietor is not related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or

the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

*[If a partnership or cooperative:]* None of the officers and members of *[Name of Bidder]* is related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

*[If a corporation or joint venture:]* None of the officers, directors, and controlling stockholders of *[Name of Bidder]* is related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

7. *[Name of Bidder]* complies with existing labor laws and standards; and
8. *[Name of Bidder]* is aware of and has undertaken the responsibilities as a Bidder in compliance with the Philippine Bidding Documents, which includes:
  - a. Carefully examining all of the Bidding Documents;
  - b. Acknowledging all conditions, local or otherwise, affecting the implementation of the Contract;
  - c. Making an estimate of the facilities available and needed for the contract to be bid, if any; and
  - d. Inquiring or securing Supplemental/Bid Bulletin(s) issued for the *[Name of the Project]*.
9. *[Name of Bidder]* did not give or pay directly or indirectly, any commission, amount, fee, or any form of consideration, pecuniary or otherwise, to any person or official, personnel or representative of the government in relation to any procurement project or activity.
- 10. In case advance payment was made or given, failure to perform or deliver any of the obligations and undertakings in the contract shall be sufficient grounds to constitute criminal liability for Swindling (Estafa) or the commission of fraud with unfaithfulness or abuse of confidence through misappropriating or converting any payment received by a person or entity under an obligation involving the duty to deliver certain goods or services, to the prejudice of the public and the government of the Philippines pursuant to Article 315 of Act No. 3815 s. 1930, as amended, or the Revised Penal Code.**

IN WITNESS WHEREOF, I have hereunto set my hand this \_\_\_ day of \_\_\_, 20\_\_ at \_\_\_\_\_, Philippines.

*[Insert NAME OF BIDDER OR ITS AUTHORIZED REPRESENTATIVE]*

*[Insert signatory's legal capacity]*

Affiant

**[Jurat]**

*[Format shall be based on the latest Rules on Notarial Practice]*

**Performance Securing Declaration (Revised)**

*[if used as an alternative performance security but it is not required to be submitted with the Bid, as it shall be submitted within ten (10) days after receiving the Notice of Award]*

REPUBLIC OF THE PHILIPPINES)  
CITY OF \_\_\_\_\_ ) S.S.

**PERFORMANCE SECURING DECLARATION**

Invitation to Bid: [Insert Reference Number indicated in the Bidding Documents]  
To: [Insert name and address of the Procuring Entity]

I/We, the undersigned, declare that:

1. I/We understand that, according to your conditions, to guarantee the faithful performance by the supplier/distributor/manufacturer/contractor/consultant of its obligations under the Contract, I/we shall submit a Performance Securing Declaration within a maximum period of ten (10) calendar days from the receipt of the Notice of Award prior to the signing of the Contract.
2. I/We accept that: I/we will be automatically disqualified from bidding for any procurement contract with any procuring entity for a period of one (1) year for the first offense, or two (2) years **for the second offense**, upon receipt of your Blacklisting Order if I/We have violated my/our obligations under the Contract;
3. I/We understand that this Performance Securing Declaration shall cease to be valid upon:
  - a. issuance by the Procuring Entity of the Certificate of Final Acceptance, subject to the following conditions:
    - i. Procuring Entity has no claims filed against the contract awardee;
    - ii. It has no claims for labor and materials filed against the contractor; and
    - iii. Other terms of the contract; or
  - b. replacement by the winning bidder of the submitted PSD with a performance security in any of the prescribed forms under Section 39.2 of the 2016 revised IRR of RA No. 9184 as required by the end-user.

**IN WITNESS WHEREOF**, I/We have hereunto set my/our hand/s this \_\_\_\_ day of [month] [year] at [place of execution].

*[Insert NAME OF BIDDER OR ITS AUTHORIZED REPRESENTATIVE]*

*[Insert signatory's legal capacity]*

Affiant

**[Jurat]**

*[Format shall be based on the latest Rules on Notarial Practice]*

